

Certificate of Analysis

Reference Material & Standard

Product Name:		Catalogue Number:	CCEM501992
Synonyms:	N/A	Lot Number:	1008-SA-7289
CAS Number:	872398-75-9	Expiry Date:	Oct 07, 28
Molecular Formula:	N/A	Storage:	2°C ~ 8°C Protected from light and Sealed
Molecular Weight:	N/A		

Warning:



Chemical Structure: N/A

Test Items	Method	Results
Appearance	Visual inspection	Solid
Identification:	LC&HNMR	Confirmed
Purity:	HPLC	96.831%
Uncertainty		±2.0%

Comment:

The uncertainty values are the expanded uncertainty and represent an estimated standard deviation equal to the positive square root of the total variation of the uncertainty of components. A normal distribution is assumed and a coverage factor of K=2 is chosen using approximately a 95% confidence level.

Signature:

Prepared by:

Lisa

Reviewed by:

Yuki

Approved by General Manager:

David

Date: Oct 08, 24

Date: Oct 08, 24

Date: Oct 08, 24



Add: 3rd Floor, Building B, No.179, Guangpu Rd East, Huangpu District, Guangzhou, China

Tel: +86-20-81960175

Email: supports@cato-chem.com

Web: <https://www.cato-chem.com>

Page 1 of 6



Directions:

- 1. Intended use: This product is suitable for identification and quantification of the analyte(s), instrument calibration, method development and quality control in analytical and R&D applications. It is not intended for administration to humans or animals, and shall not be used to diagnose, treat or cure any diseases.*
- 2. Minimum sample size: 1mg is recommended as the minimum sample size.*
- 3. Metrological traceability: This product meets the requirements of ISO 17034 and is traceable to NMI through an unbroken chain of comparisons.*
- 4. Instructions for handling and correct use: Handling in a well-ventilated place.*

Document Version:

Revision No.	Date	Reason for Revision
01	Oct 08, 24	First release.

