


Certificate of Analysis

Reference Material & Standard

Product Name:	Chloriodomethane (stabilized with Copper chip)	Catalogue Number:	CCHM705878
Synonyms:	N/A	Lot Number:	0113-RG-0155
CAS Number:	593-71-5	Expiry Date:	Jan 14, 2029
Molecular Formula:	CH ₂ ClI	Storage:	Room Temp Protected from light and Sealed
Molecular Weight:	176.38	Warning:	

Chemical Structure:



Test Items	Method	Results
Appearance	Visual inspection	Colourless Liquid
Identification	GCMS&HNMR	Confirmed
Purity	GC	99.382%
Water Content	Karl Fischer	0.0758%
Assay		99.3%
Uncertainty		±2.0%

Comment:

1. The calculation of the 100% method follows the formula (calculated on the non-dried basis):

$$\text{Assay (\%)} = (100\% - \text{Water Content}) * \text{Purity GC (\%)} / 100\%$$

Water Content is considered as absolute effect, GC purity is considered as relative effect.

2. The uncertainty values are the expanded uncertainty and represent an estimated standard deviation equal to the positive square root of the total variation of the uncertainty of components. A normal distribution is assumed and a coverage factor of K=2 is chosen using approximately a 95% confidence level.

Signature:

Prepared by:

Nice

Reviewed by:

Lisa

Approved by General Manager:

David

Date: Jan 15, 2025

Date: Jan 15, 2025

Date: Jan 15, 2025

