


Certificate of Analysis

Reference Material & Standard

Product Name: Aroclor 1254
Appearance: Colourless Liquid
Solvent: Methanol
Expiry Date: Feb 03, 2028
Chemical Structure:



Warning:



NO.	Component	CAS No.	Certified Concentration Value (µg/mL)	Uncertainty (µg/mL)
1	Aroclor 1254	11097-69-1	200.4	6.1

Comment:

The uncertainty values are the expanded uncertainty and represent an estimated standard deviation equal to the positive square root of the total variation of the uncertainty of components. A normal distribution is assumed and a coverage factor of $K=2$ is chosen using approximately a 95% confidence level.

Signature:

Prepared by: 古雪

Reviewed by:

Xenia

Aproved by General Manager:

David

Date: Feb 04, 2023

Date: Jun 08, 2023

Date: Jun 08, 2023

